



## Post 12 lead ECG:

1. locate V4R position:
  - 5<sup>th</sup> intercostal space
  - right side mid-clavicular line
2. move **V1** lead to the **V4R** position
3. locate V8 position:
  - in line with anterior 5<sup>th</sup> intercostal space
  - mid scapular line
4. move **V5** lead to **V8** position
5. locate V9 position:
  - in line with V8 position
  - left paraspinal border
6. move **V6** lead to **V9** position



7. Acquire 15 lead ECG: press '12 lead ECG' and 'start acquire' (a new 'acquire' must be made, or the old 12 lead capture will print)
8. When the monitor prints the new ECG, re-label the three altered leads on the ECG strip (i.e. change V1 to V4R, V5 to V8 and V6 to V9)